

NATA Accredited Laboratory

National Association of Testing Authorities, Australia
(ABN 59 004 379 748)

has accredited

ACO Polycrete Pty Ltd Testing Laboratory

following demonstration of its technical competence
to operate in accordance with

ISO/IEC 17025

This facility is accredited in the field of

Mechanical Testing

for the tests shown on the *Scope of Accreditation* issued by NATA



A J Russell
Chief Executive

Date of accreditation: 26 June 2006
Accreditation number: 15193



**WORLD RECOGNISED
ACCREDITATION**

NATA is Australia's government-endorsed laboratory accreditor, and a leader in accreditation internationally. NATA is a signatory to the international mutual recognition arrangements of the International Laboratory Accreditation Cooperation (ILAC) and the Asia Pacific Laboratory Accreditation Cooperation (APLAC).



Scope of Accreditation

ACCREDITATION NO: 15193

ACO Polycrete Pty Ltd

Testing Laboratory
134-140 Old Bathurst Road
EMU PLAINS NSW 2750

CONTACT: Mr D Chan
PHONE: (02) 4747 4000 FAX: (02) 4747 4060 MOBILE:
EMAIL: dchan@acoaus.com.au
WEB SITE: www.acoaus.com.au

FACILITIES: Normally not available for public testing

This facility complies with the requirements of ISO/IEC 17025:2005

13.01 Metals and metal products

.51 Compression, transverse and shear tests
Compression and deflection tests on metal access covers and metal grates
by the methods of
AS 3996 Appendix C, EN 124 Clause 8 and EN 1433 Clause 9.1
in the range 2 to 1000 kN

13.40 Other materials

Compression and deflection tests on composite access covers, composite inspection covers and composite
grates
by the methods of
AS 3996 Appendix C, EN 124 Clause 8 and EN 1433 Clause 9.1
in the range 2 to 1000 kN

Accreditation No: 15193
(Scope Last Changed 08/08/06)
