



# NATA ACCREDITED LABORATORY

National Association of Testing Authorities, Australia

(ABN 59 004 379 748)

has accredited

## ACO Polycrete Pty Ltd Testing Laboratory

following demonstration of its technical competence to operate in accordance with

### ISO/IEC 17025

This facility is accredited in the field of

### MECHANICAL TESTING

for the tests, calibrations and measurements shown on the *Scope of Accreditation*  
issued by NATA

**Jennifer Evans**  
Chief Executive Officer

Date of issue: 16 February 2016

Date of accreditation: 26 June 2006

Accreditation number: 15193

# Scope of Accreditation



**ACCREDITATION NO: 15193**

**ACO Polycrete Pty Ltd**

Testing Laboratory  
134-140 Old Bathurst Road  
EMU PLAINS NSW 2750

CONTACT: Mr D Chan  
PHONE: (02) 4747 4000 FAX: (02) 4747 4060 MOBILE: 0433 997 097  
EMAIL: [dchan@acoaus.com.au](mailto:dchan@acoaus.com.au)  
WEB SITE: [www.acoaus.com.au](http://www.acoaus.com.au)

FACILITIES: Normally not available for public testing

---

**This facility complies with the requirements of ISO/IEC 17025:2005**

**13.01 Metals and metal products**

.51 Compression, transverse and shear tests  
Compression and deflection tests on metal access covers and metal grates  
by the methods of  
AS 3996 Appendix C, EN 124-1 Annexes A and B and EN 1433 Clause 9.1  
in the range 2 to 1000 kN

**13.99 Other materials and products**

Compression and deflection tests on composite access covers, composite inspection  
covers and composite grates  
by the methods of  
AS 3996 Appendix C, EN 124-1 Annexes A and B and EN 1433 Clause 9.1  
in the range 2 to 1000 kN

Accreditation No: 15193  
(Scope Last Changed 16/02/16 )

---